



CERTIFICATE OF Conformity



EC Council Directive 2004/108/EC

Electromagnetic Comppatibility

Registration No : AT1207674E

Raport No : 201207734E

Firma:

IQ WELL HANDELSAGENTUR & HANDEL mit MEDIZINPRODUKTE e.U.

Mag. Danuta Gugglberger KROTTENSEESTR.45 A - 4810 GMUNDEN

www.iq-well.at d.gug@gmx.at iqwell@gmx.at

UID Nr. ATU22005005 Steuer-Nr. 53 111/4924 EORI-Nr. ATEOS1000083056

Produuct : IQ-WELL Bio-SCANNER

Identification:

Trade Mark: GY

Rating : DC 4,7 – 5,25 V 200mA Max.

Test : EN 55011:2009

Standards : EN 60601-1-2 : 2007

The certificate of conformity is based on an evaluation of a sample of the above-mentioned product. Technical report and documentation are at the applicant`s disposal. This is to certify that the tested sample is in conformity with all provisions of Annex III of Council Directive 2004/108/EC, in its latest amended version, reverred to EMC Direktive. The certivicate does not imply assessment of the production and does not permit the use of Labs`s logo. The applicant of the certificate is authorized to use this certificate in connection with EC declaration of conformity to Article 10.1 of the Directive.

Gmunden Date 08.11.2017


Mag. Danuta Gugglberger
Handelsagentur & Handel mit
Medizinprodukten
Mag. D. GUGGLBERGER
KROTTENSEESTR.18 4810 GMUNDEN
A 0676-501 5272 d.gug@gmx.at
UID Nr. ATU22005005